

WIKAL Alexander Wiegand SE & Co. KG  
 PF 1180 • 63908 Klingenberg • Germany

Engemann u. Co Internationale Spedition GmbH  
 For Alfons Köster / in order Delta GmbH  
 Otto-Hahn-Str. 29  
 40721 Hilden  
 Germany

*Order address:*  
 DELTA TECH GmbH  
 Farshad Shoarian  
 ALSTER TOWER  
 Lübecker Straße 1  
 22087 Hamburg  
 Germany

## Delivery note

# DN00730438

18/12/2024

Customer no.	43327028	Sales order no.	S110232513
Order no.	DELTA-WIKA-2024-PO-200	Contact	Deepali Lokare
Order date	03/10/2024	E-Mail	Deepali.Lokare@wika.com
Contact	Farshad Shoarian		
VAT-No.	DE4871404580		
Invoice address	DELTA TECH GmbH ALSTER TOWER Lübecker Straße 1 22087 Hamburg Germany	Gross/net weight	0,00 / 0,00 kg
		Mode of delivery	Storage Hall 6
		Terms of delivery	FCA - Free CARRIER

\*\*\*\*\*  
 \*\*\*\*\*  
 Note: For tags LS-RU0001A-01, LS-RU0001B-01, we have not offered the Level switch, as it is not clear from the specifications, which type of the float is required.  
 \*\*\*\*\*  
 \*\*\*\*\*

Line	Qty Unit	Item
001	2,00 Pcs	14728736 233.30.100 -1/+24 barg 1/2 NPT lower mount Measuring System 316L Ingress protection: IP 65 Case material: Stainless steel 316Ti class 1.0 Window: laminated safety glass adjustable pointer Case filling fluid: glycerine Permissible medium temperature: + 100 °C Permissible ambient temperature: -20 ... +60 °C Pressure Gauge Standard: International (standard Europe) with Serial number Test report 2.2 per EN 10204 Certificate Language: English and German

WIKAL Alexander Wiegand SE & Co. KG  
 Alexander-Wiegand-Straße 30  
 63911 Klingenberg  
 Germany  
 Phone +49 9372 132-0  
 Fax +49 9372 132-406  
 info@wika.de  
 www.wika.de  
 WEEE-Reg-Nr. DE 92770372  
 VAT.No.: DE131878642

Limited partnership company:  
 Place of business Klingenberg  
 Local court Aschaffenburg HRA 1819  
 General partner:  
 WIKAL International SE – Pöb Klingenberg  
 Local court Aschaffenburg HRB 10505  
 Chairman of the board: Alexander Wiegand  
 Chairman of the supervisory board: Prof. Dr. Roderich C. Thümmel

Bank account:  
 Commerzbank Aschaffenburg  
 BIC: COBADEFF795  
 IBAN DE57 7954 0049 0104 1177 00  
 Sparkasse Aschaffenburg Miltenberg  
 BIC: BYLADEM1ASA  
 IBAN: DE95 7955 0000 0430 1062 94  
 Sydbank (DKK Payments)  
 BIC: SYBKDK22  
 IBAN: DK02 8072 0001 0403 60

**Delivery note**                      **DN00730438**  
 43327028 DELTA TECH GmbH, Hamburg

Page 2 / 10  
 18/12/2024

**Line**                      **Qty Unit**                      **Item**

Certificate for: pressure gauge  
 Formal certificate: Manufactured according to state-of-the-art  
 Inspection certificate 3.1 per EN 10204  
 Certificate type: Inspection certificate acc. to DIN EN 10204 3.1  
 Certificate Language: English and German  
 Certificate for: pressure gauge  
 Accuracy: Calibration values listed  
 Material certificate: Wetted parts material (metal contents) with mill certificate incl. chemical analysis  
 Pressure and stability test: 1.3 x full scale range; duration 5 minutes  
 Label: Standard neutral, without / with Tag-number  
 mechanical pressure gauge

TAG-1 (SS tag plate)  
 PG-RU0001A-01; PG-RU0001B-01  
 Customs tariff number: 90262040  
 Country of origin: PL  
 preferential: Yes

002                      2,00 Pcs                      14728750

IPT-20, -1...+40 bar gauge, Thread 1/2 NPT PN1000, ASME B1.20.1 / 316L Sealing: without, -40 ... +105°C, 4...20 mA/HART , ATEX II 1G Ex ia IIC T6...T1 Ga  
 Accuracy: 0.1% , Alu single compact/ IP66/IP67; NEMA 4X, display on top  
 M20x1.5 cable gland brass 5-9mm  
 Scaling: -1.00 ... +10.30 bar ,  
 Mounting bracket for wall and pipe mounting (stainless steel) (for 2" pipe)  
 Test report 2.2 per EN 10204  
 Certificate Language: English and German  
 Certificate for: process pressure transmitter  
 Formal certificate: Manufactured according to state-of-the-art  
 Inspection certificate 3.1 per EN 10204  
 Certificate type: Inspection certificate acc. to DIN EN 10204 3.1  
 Certificate Language: English and German  
 Certificate for: process pressure transmitter  
 Accuracy: Calibration values listed  
 Material certificate: Wetted parts material (metal contents) with mill certificate incl. chemical analysis  
 Pressure and stability test: 1.3 x full scale range; duration 5 minutes

Deviation:  
 1) Required element material stainless steel and Hastelloy offered stainless steel, please check  
 2) IPT-20 suitable temp. is -40 DegC  
 3) Required Wetted O-Rings Material : Teflon-coated 316L: not offered  
 4) Housing material is: Single chamber case, powder-coated aluminium  
 5) Accuracy: Calibration values listed : this certificate is 5- point calibration  
 6) Quotated pressure transmitter process connection is 1/2" NPT M.

TAG-1 (SS tag plate)  
 PIT-RU0001A-01; PIT-RU0001B-01  
 Customs tariff number: 90262020  
 Country of origin: DE  
 preferential: Yes

003                      8,00 Pcs                      14728751

WIKAL Alexander Wiegand SE & Co. KG  
 Alexander-Wiegand-Straße 30  
 63911 Klingenberg  
 Germany  
 Phone +49 9372 132-0  
 Fax +49 9372 132-406  
 info@wika.de  
 www.wika.de  
 WEEE-Reg-Nr. DE 92770372  
 VAT.No.: DE131878642

Limited partnership company:  
 Place of business Klingenberg  
 Local court Aschaffenburg HRA 1819  
 General partner:  
 WIKAL International SE – Pob Klingenberg  
 Local court Aschaffenburg HRB 10505  
 Chairman of the board: Alexander Wiegand  
 Chairman of the supervisory board: Prof. Dr. Roderich C. Thümmel

Bank account:  
 Commerzbank Aschaffenburg  
 BIC: COBADEFF795  
 IBAN DE57 7954 0049 0104 1177 00  
 Sparkasse Aschaffenburg Miltenberg  
 BIC: BYLADEM1ASA  
 IBAN: DE95 7955 0000 0430 1062 94  
 Sydbank (DKK Payments)  
 BIC: SYBKDK22  
 IBAN: DK02 8072 0001 0403 60

**Delivery note**                      **DN00730438**  
 43327028 DELTA TECH GmbH, Hamburg

Page 3 / 10  
 18/12/2024

Line	Qty Unit	Item
		<p>IPT-20, -1...+40 bar gauge, Thread 1/2 NPT PN1000, ASME B1.20.1 / 316L Sealing: without, -40 ... +105°C, 4...20 mA/HART , ATEX II 1G Ex ia IIC T6...T1 Ga            Accuracy: 0.1% , Alu single compact/ IP66/IP67; NEMA 4X, display on top            M20x1.5 cable gland brass 5-9mm            Scaling: -1.00 ... +27.60 bar ,            Mounting bracket for wall and pipe mounting (stainless steel) (for 2" pipe)            Test report 2.2 per EN 10204            Certificate Language: English and German            Certificate for: process pressure transmitter            Formal certificate: Manufactured according to state-of-the-art            Inspection certificate 3.1 per EN 10204            Certificate type: Inspection certificate acc. to DIN EN 10204 3.1            Certificate Language: English and German            Certificate for: process pressure transmitter            Accuracy: Calibration values listed            Material certificate: Wetted parts material (metal contents) with mill certificate incl. chemical analysis            Pressure and stability test: 1.3 x full scale range; duration 5 minutes</p> <p>TAG-1 (SS tag plate)            PIT-RU0001A-02; PIT-RU0001A-03; PIT-RU0001A-05; PIT-RU0001A-06; PIT-RU0001B-02; PIT-RU0001B-03; PIT-RU0001B-05; PIT-RU0001B-06            Customs tariff number: 90262020            Country of origin: DE            preferential: No</p>
004	2,00 Pcs	14728751 <p>IPT-20, -1...+40 bar gauge, Thread 1/2 NPT PN1000, ASME B1.20.1 / 316L Sealing: without, -40 ... +105°C, 4...20 mA/HART , ATEX II 1G Ex ia IIC T6...T1 Ga            Accuracy: 0.1% , Alu single compact/ IP66/IP67; NEMA 4X, display on top            M20x1.5 cable gland brass 5-9mm            Scaling: -1.00 ... +27.60 bar ,            Mounting bracket for wall and pipe mounting (stainless steel) (for 2" pipe)            Test report 2.2 per EN 10204            Certificate Language: English and German            Certificate for: process pressure transmitter            Formal certificate: Manufactured according to state-of-the-art            Inspection certificate 3.1 per EN 10204            Certificate type: Inspection certificate acc. to DIN EN 10204 3.1            Certificate Language: English and German            Certificate for: process pressure transmitter            Accuracy: Calibration values listed            Material certificate: Wetted parts material (metal contents) with mill certificate incl. chemical analysis            Pressure and stability test: 1.3 x full scale range; duration 5 minutes</p> <p>TAG-1 (SS tag plate)            PIT-RU0001A-04; PIT-RU0001B-04            Customs tariff number: 90262020            Country of origin: DE            preferential: Yes</p>
005	2,00 Pcs	5181.63

WIKAL Alexander Wiegand SE & Co. KG  
 Alexander-Wiegand-Straße 30  
 63911 Klingenberg  
 Germany  
 Phone +49 9372 132-0  
 Fax +49 9372 132-406  
 info@wika.de  
 www.wika.de  
 WEEE-Reg-Nr. DE 92770372  
 VAT.No.: DE131878642

Limited partnership company:  
 Place of business Klingenberg  
 Local court Aschaffenburg HRA 1819  
 General partner:  
 WIKAL International SE – Pob Klingenberg  
 Local court Aschaffenburg HRB 10505  
 Chairman of the board: Alexander Wiegand  
 Chairman of the supervisory board: Prof. Dr. Roderich C. Thümmel

Bank account:  
 Commerzbank Aschaffenburg  
 BIC: COBADEFF795  
 IBAN DE57 7954 0049 0104 1177 00  
 Sparkasse Aschaffenburg Miltenberg  
 BIC: BYLADEM1ASA  
 IBAN: DE95 7955 0000 0430 1062 94  
 Sydbank (DKK Payments)  
 BIC: SYBKDK22  
 IBAN: DK02 8072 0001 0403 60

**Delivery note**                      **DN00730438**  
43327028 DELTA TECH GmbH, Hamburg

Page 4 / 10  
18/12/2024

Line	Qty	Unit	Item
			TR12-B, Ex i, 1xPt100, 3-w, Cl.A, -100...+450°C, Ø 6 mm, T32 ATEX, Ex i, Zone 1 gas [ II 2G Ex ia IIC T1...T6 Gb ]  Head / housing: 5/6000 (Alu), DIH50, 2x M20 x 1.5 Cable entry A: Thread (female) Cable entry B: Closing plug Transmitter: T32, Digital-transmitter, HART®, 4...20 mA, universal programmable Customer specified  Nipple-Union-Nipple Connection to housing / head: Male thread M20 x 1,5, with adjustable lock nut Connection to thermowell: Male thread M20 x 1,5 Neck length [N/MH]: 150 mm (approx. 6.0 inch)  Insertion length [A/L1/U2]: 120 mm; Diameter measuring insert: 6 mm - 1.4571,  Specifications according to data sheet: TE 60.17 Measuring insert length (spare) [J5]: 288 mm

TW15-H 1/2 BSPT M20x1,5 female U=70 mm B=0.260 inch (6.6 mm) SS316/316L  
Process thread [P or E]: 1/2 BSPT  
Connection to thermometer [N]: M20x1,5 female  
Thermowell style: Tapered  
Head length [H]: 50 mm  
Insertion length [U]: 70 mm  
Root diameter [Q]: 16 mm  
Tip diameter [V or F3]: 13 mm  
Bore size [B or d1]: 0.260 inch (6.6 mm)  
Bore depth / stem length (US) [A]: 113 mm  
Specifications according to data sheet: TW 95.15

Test certificate acc. to EN 10204 2.2  
Certificate Language: English and German  
Thermowell / protection tube  
Formal certificate: Manufactured according to state-of-the-art  
Inspection certificate acc. to DIN EN 10204 3.1  
Certificate Language: English and German  
Thermowell / protection tube  
Material certificate: Wetted parts material (metal contents) with mill certificate incl. chemical analysis  
Pressure and stability test: Hydrostatic, internal pressure test - 500 bar (7500 psig) - 3 min.  
Inspection certificate acc. to DIN EN 10204 3.1  
Certificate Language: English and German  
Measuring insert / probe with transmitter  
Accuracy: 3 test points (0°C / 50°C / 100°C) for RTD  
Digital temperature transmitter T32  
with HART-protocol  
Specifications according to data sheet: TE 32.04  
Output signal: 4 ... 20 mA, 2-wire-design  
Input: configurable by means of software  
Measuring range: configurable by means of software  
Model: T32.1S, head mounting

WIKA Alexander Wiegand SE & Co. KG  
Alexander-Wiegand-Straße 30  
63911 Klingenberg  
Germany  
Phone +49 9372 132-0  
Fax +49 9372 132-406  
info@wika.de  
www.wika.de  
WEEE-Reg-Nr. DE 92770372  
VAT.No.: DE131878642

Limited partnership company:  
Place of business Klingenberg  
Local court Aschaffenburg HRA 1819  
General partner:  
WIKA International SE – Pob Klingenberg  
Local court Aschaffenburg HRB 10505  
Chairman of the board: Alexander Wiegand  
Chairman of the supervisory board: Prof. Dr. Roderich C. Thümmel

Bank account:  
Commerzbank Aschaffenburg  
BIC: COBADEFF795  
IBAN DE57 7954 0049 0104 1177 00  
Sparkasse Aschaffenburg Miltenberg  
BIC: BYLADEM1ASA  
IBAN DE95 7955 0000 0430 1062 94  
Sydbank (DKK Payments)  
BIC: SYBKDK22  
IBAN: DK02 8072 0001 0403 60

**Delivery note**                      **DN00730438**  
 43327028 DELTA TECH GmbH, Hamburg

Page 5 / 10  
 18/12/2024

**Line**                      **Qty Unit**                      **Item**

Explosion protection: Ex ia, for ATEX/IEC/CSA/FM/EAC/KOSHA/NEPSI/INMETRO/PESO  
 SIL-Statement: not for SIL applications  
 HART revision: HART® 5  
 Configuration: customer's specification  
 Permissible ambient temperature [Transmitter]: -40 °C ... +85 °C  
 Certificate: calibration values listed  
 Transmitter configuration:  
 Input: resistance thermometer Pt100  
 Sensors: Single-sensor  
 Sensor-calculation-functionality: Sensor 1, Sensor 2 not available  
 Sensor-Drift-Detection: not active  
 Input configuration: 3-wire  
 Unit: °C  
 Measuring range: -50 °C ... +100 °C  
 Output signal: 4 ... 20 mA, linear to temperature  
 Lower output limit: to NAMUR (3,8 mA)  
 Upper output limit: to NAMUR (20,5 mA)  
 Current value for signalling down scale: 3,5 mA (NAMUR down scale)  
 Current value for signalling up scale: 21,5 mA (NAMUR up scale)  
 Signalling of error: down scale (uniform for all kinds of errors)  
 Damping: off  
 Inspection certificate acc. to DIN EN 10204 3.1  
 Certificate Language: English and German  
 Transmitter  
 Accuracy: 3 test points (start / middle / end of transmitter range)

Deviation:  
 1) Offered Neck length [N/MH]: 150 mm (approx. 6.0 inch) instead of 50 mm, please check and confirm.  
 2) Clarification: 70 mm is a thermowell U length and 120 mm is a sensor length (i.e. U length + H Length= sensor A length).  
 3) Ingress protection : IP66.  
 4) Housing material is Aluminum painted

TAG-1 (SS tag plate)  
 TIT-RU0001A-01; TIT-RU0001B-01  
 Customs tariff number: 90251900  
 Country of origin: DE  
 preferential: Yes

006                      2,00 Pcs                      5181.63  
 TR12-B, Ex i, 1xPt100, 3-w, Cl.A, -100...+450°C, Ø 6 mm, T32  
 ATEX, Ex i, Zone 1 gas [ II 2G Ex ia IIC T1...T6 Gb ]

Head / housing: 5/6000 (Alu), DIH50, 2x M20 x 1.5  
 Cable entry A: Thread (female)  
 Cable entry B: Closing plug  
 Transmitter: T32, Digital-transmitter, HART®, 4...20 mA, universal programmable  
 Customer specified

Nipple-Union-Nipple  
 Connection to housing / head: Male thread M20 x 1,5, with adjustable lock nut  
 Connection to thermowell: Male thread M20 x 1,5

**Delivery note**                      **DN00730438**  
 43327028 DELTA TECH GmbH, Hamburg

Page 6 / 10  
 18/12/2024

**Line**                      **Qty Unit**                      **Item**

Neck length [N/MH]: 150 mm (approx. 6.0 inch)

Insertion length [A/L1/U2]: 90 mm;  
 Diameter measuring insert: 6 mm - 1.4571,

Specifications according to data sheet: TE 60.17  
 Measuring insert length (spare) [I5]: 258 mm

TW15-H 1/2 BSPT M20x1,5 female U=40 mm B=0.260 inch (6.6 mm) SS316/316L  
 Process thread [P or E]: 1/2 BSPT  
 Connection to thermometer [N]: M20x1,5 female  
 Thermowell style: Tapered  
 Head length [H]: 50 mm  
 Insertion length [U]: 40 mm  
 Root diameter [Q]: 16 mm  
 Tip diameter [V or F3]: 13 mm  
 Bore size [B or d1]: 0.260 inch (6.6 mm)  
 Bore depth / stem length (US) [A]: 83 mm  
 Specifications according to data sheet: TW 95.15

Test certificate acc. to EN 10204 2.2  
 Certificate Language: English and German  
 Thermowell / protection tube  
 Formal certificate: Manufactured according to state-of-the-art  
 Inspection certificate acc. to DIN EN 10204 3.1  
 Certificate Language: English and German  
 Thermowell / protection tube  
 Material certificate: Wetted parts material (metal contents) with mill certificate incl. chemical analysis  
 Pressure and stability test: Hydrostatic, internal pressure test - 500 bar (7500 psig) - 3 min.  
 Inspection certificate acc. to DIN EN 10204 3.1  
 Certificate Language: English and German  
 Measuring insert / probe with transmitter  
 Accuracy: 3 test points (0°C / 50°C / 100°C) for RTD  
 Digital temperature transmitter T32  
 with HART-protocol  
 Specifications according to data sheet: TE 32.04  
 Output signal: 4 ... 20 mA, 2-wire-design  
 Input: configurable by means of software  
 Measuring range: configurable by means of software  
 Model: T32.1S, head mounting  
 Explosion protection: Ex ia, for ATEX/IEC/CSA/FM/EAC/KOSHA/NEPSI/INMETRO/PESO  
 SIL-Statement: not for SIL applications  
 HART revision: HART® 5  
 Configuration: customer's specification  
 Permissible ambient temperature [Transmitter]: -40 °C ... +85 °C  
 Certificate: calibration values listed  
 Transmitter configuration:  
 Input: resistance thermometer Pt100  
 Sensors: Single-sensor  
 Sensor-calculation-functionality: Sensor 1, Sensor 2 not available  
 Sensor-Drift-Detection: not active  
 Input configuration: 3-wire

**Delivery note**                      **DN00730438**  
 43327028 DELTA TECH GmbH, Hamburg

Page 7 / 10  
 18/12/2024

**Line**                      **Qty Unit**                      **Item**

Unit: °C  
 Measuring range: -50 °C ... +100 °C  
 Output signal: 4 ... 20 mA, linear to temperature  
 Lower output limit: to NAMUR (3,8 mA)  
 Upper output limit: to NAMUR (20,5 mA)  
 Current value for signalling down scale: 3,5 mA (NAMUR down scale)  
 Current value for signalling up scale: 21,5 mA (NAMUR up scale)  
 Signalling of error: down scale (uniform for all kinds of errors)  
 Damping: off  
 Inspection certificate acc. to DIN EN 10204 3.1  
 Certificate Language: English and German  
 Transmitter  
 Accuracy: 3 test points (start / middle / end of transmitter range)

Deviation:  
 1) Offered Neck length [N/MH]: 150 mm (approx. 6.0 inch) instead of 50 mm, please check and confirm.  
 2) Clarification: 40 mm is a thermowell U length and 90 mm is a sensor length (i.e. U length + H Length= sensor A length).  
 3) Ingress protection : IP66.  
 5) Housing material is Aluminum painted

TAG-1 (SS tag plate)  
 TIT-RU0001A-02; TIT-RU0001B-02  
 Customs tariff number: 90251900  
 Country of origin: DE  
 preferential: Yes

007                      2,00 Pcs                      5181.63  
 TR12-B, Ex i, 1xPt100, 3-w, Cl.A, -100...+450°C, Ø 6 mm, T32  
 ATEX, Ex i, Zone 1 gas [ II 2G Ex ia IIC T1...T6 Gb ]

Head / housing: 5/6000 (Alu), DIH50, 2x M20 x 1.5  
 Cable entry A: Thread (female)  
 Cable entry B: Closing plug  
 Transmitter: T32, Digital-transmitter, HART®, 4...20 mA, universal programmable  
 Customer specified

Nipple-Union-Nipple  
 Connection to housing / head: Male thread M20 x 1,5, with adjustable lock nut  
 Connection to thermowell: Male thread M20 x 1,5  
 Neck length [N/MH]: 150 mm (approx. 6.0 inch)

Insertion length [A/L1/U2]: 250 mm;  
 Diameter measuring insert: 6 mm - 1.4571,

Specifications according to data sheet: TE 60.17  
 Measuring insert length (spare) [I5]: 418 mm

TW15-H 1/2 BSPT M20x1,5 female U=200 mm B=0.260 inch (6.6 mm) SS316/316L  
 Process thread [P or E]: 1/2 BSPT  
 Connection to thermometer [N]: M20x1,5 female  
 Thermowell style: Tapered

**Delivery note**                      **DN00730438**  
 43327028 DELTA TECH GmbH, Hamburg

Page 8 / 10  
 18/12/2024

**Line**                      **Qty Unit**                      **Item**

Head length [H]: 50 mm  
 Insertion length [U]: 200 mm  
 Root diameter [Q]: 16 mm  
 Tip diameter [V or F3]: 13 mm  
 Bore size [B or d1]: 0.260 inch (6.6 mm)  
 Bore depth / stem length (US) [A]: 243 mm  
 Specifications according to data sheet: TW 95.15

Test certificate acc. to EN 10204 2.2  
 Certificate Language: English and German  
 Thermowell / protection tube  
 Formal certificate: Manufactured according to state-of-the-art  
 Inspection certificate acc. to DIN EN 10204 3.1  
 Certificate Language: English and German  
 Thermowell / protection tube  
 Material certificate: Wetted parts material (metal contents) with mill certificate incl. chemical analysis  
 Pressure and stability test: Hydrostatic, internal pressure test - 500 bar (7500 psig) - 3 min.  
 Inspection certificate acc. to DIN EN 10204 3.1  
 Certificate Language: English and German  
 Measuring insert / probe with transmitter  
 Accuracy: 3 test points (0°C / 50°C / 100°C) for RTD  
 Digital temperature transmitter T32  
 with HART-protocol  
 Specifications according to data sheet: TE 32.04  
 Output signal: 4 ... 20 mA, 2-wire-design  
 Input: configurable by means of software  
 Measuring range: configurable by means of software  
 Model: T32.1S, head mounting  
 Explosion protection: Ex ia, for ATEX/IEC/CSA/FM/EAC/KOSHA/NEPSI/INMETRO/PESO  
 SIL-Statement: not for SIL applications  
 HART revision: HART® 5  
 Configuration: customer's specification  
 Permissible ambient temperature [Transmitter]: -40 °C ... +85 °C  
 Certificate: calibration values listed  
 Transmitter configuration:  
 Input: resistance thermometer Pt100  
 Sensors: Single-sensor  
 Sensor-calculation-functionality: Sensor 1, Sensor 2 not available  
 Sensor-Drift-Detection: not active  
 Input configuration: 3-wire  
 Unit: °C  
 Measuring range: -50 °C ... +100 °C  
 Output signal: 4 ... 20 mA, linear to temperature  
 Lower output limit: to NAMUR (3,8 mA)  
 Upper output limit: to NAMUR (20,5 mA)  
 Current value for signalling down scale: 3,5 mA (NAMUR down scale)  
 Current value for signalling up scale: 21,5 mA (NAMUR up scale)  
 Signalling of error: down scale (uniform for all kinds of errors)  
 Damping: off  
 Inspection certificate acc. to DIN EN 10204 3.1  
 Certificate Language: English and German  
 Transmitter  
 Accuracy: 3 test points (start / middle / end of transmitter range)

WIKA Alexander Wiegand SE & Co. KG  
 Alexander-Wiegand-Straße 30  
 63911 Klingenberg  
 Germany  
 Phone +49 9372 132-0  
 Fax +49 9372 132-406  
 info@wika.de  
 www.wika.de  
 WEEE-Reg-Nr. DE 92770372  
 VAT.No.: DE131878642

Limited partnership company:  
 Place of business Klingenberg  
 Local court Aschaffenburg HRA 1819  
 General partner:  
 WIKA International SE – Pob Klingenberg  
 Local court Aschaffenburg HRB 10505  
 Chairman of the board: Alexander Wiegand  
 Chairman of the supervisory board: Prof. Dr. Roderich C. Thümmel

Bank account:  
 Commerzbank Aschaffenburg  
 BIC: COBADEFF795  
 IBAN DE57 7954 0049 0104 1177 00  
 Sparkasse Aschaffenburg Miltenberg  
 BIC: BYLADEM1ASA  
 IBAN: DE95 7955 0000 0430 1062 94  
 Sydbank (DKK Payments)  
 BIC: SYBKDK22  
 IBAN: DK02 8072 0001 0403 60

**Delivery note**                      **DN00730438**  
 43327028 DELTA TECH GmbH, Hamburg

Page 9 / 10  
 18/12/2024

Line	Qty Unit	Item
------	----------	------

Deviation:

- 1) Offered Neck length [N/MH]: 150 mm (approx. 6.0 inch) instead of 100 mm, please check and confirm.
- 2) WFC not performed due to velocity not given.
- 3) Clarification: 200 mm is a thermowell U length and 250 mm is a sensor length (i.e. U length + H Length= sensor A length).
- 4) Ingress protection : IP66.
- 5) Housing material is Aluminum painted

TAG-1 (SS tag plate)  
 TIT-RU0001A-03; TIT-RU0001B-03  
 Customs tariff number: 90251900  
 Country of origin: DE  
 preferential: Yes

011	2,00 Pcs	14731931 T38.R, ATEX/IECEX-Ex ia, customer spec. Poti Poti, +0...+100%
-----	----------	---

Specifications according to data sheet: TE 38.01  
 Poti Potentiometer  
 Total resistance (Potentiometer/FLR Sensor): 1 kOhm  
 Sensor 1, sensor 2 not available  
 Tamb: Standard -40...85°C  
 HART® 7  
 Transmitter Signalling of error: Down scale

TAG-1 (SS tag plate)  
 ZT-RU0001A-01; ZT-RU0001B-01  
 Customs tariff number: 90259000  
 Country of origin: DE  
 preferential: Yes

Information regarding WIKAL and our products as well as instruction manuals, CE Declaration of conformity and approvals can be found online at [www.wika.de](http://www.wika.de).

Thank you for the order you placed with us. We will execute it in accordance with our General Sales - and Service Conditions, which you can find at [https://www.wika.com/en-en/t\\_cs.WIKA](https://www.wika.com/en-en/t_cs.WIKA).

Suppliers' declaration for products having preferential origin status

#### DECLARATION

I, the undersigned, declare that the goods listed on this document and marked > preferential Yes < originate in the EEC / EU and satisfy the rules of origin governing preferential trade with AL, BA, CH, CL, DZ, EG, FO, IL, JO, LB, LI, MA, ME, MK, MX, NO, PS, TN, XC, XL, AG, BB, BS, BZ, DM, DO, GD, GY, JM, KN, LC, SR, TT, VC, KR, PG, FJ, IS, XS, MG, MU, SC, ZW, NI, PA, HN, CR, SV, GT, PE, CO, GE, MD, XK, UA, CI, CM, EC, BW, LS, NA, ZA, SZ.

I declare, that:

Cumulation applied with ... (name of the country/countries).  
 No cumulation applied.

I undertake to make available to the customs authorities any further supporting documents they require.

WIKAL Alexander Wiegand SE & Co. KG  
 Alexander-Wiegand-Straße 30  
 63911 Klingenberg  
 Germany  
 Phone +49 9372 132-0  
 Fax +49 9372 132-406  
[info@wika.de](mailto:info@wika.de)  
[www.wika.de](http://www.wika.de)  
 WEEE-Reg-Nr. DE 92770372  
 VAT.No.: DE131878642

Limited partnership company:  
 Place of business Klingenberg  
 Local court Aschaffenburg HRA 1819  
 General partner:  
 WIKAL International SE – Pöb Klingenberg  
 Local court Aschaffenburg HRB 10505  
 Chairman of the board: Alexander Wiegand  
 Chairman of the supervisory board: Prof. Dr. Roderich C. Thümmel

Bank account:  
 Commerzbank Aschaffenburg  
 BIC: COBADEFF795  
 IBAN DE57 7954 0049 0104 1177 00  
 Sparkasse Aschaffenburg Miltenberg  
 BIC: BYLADEM1ASA  
 IBAN: DE95 7955 0000 0430 1062 94  
 Sydbank (DKK Payments)  
 BIC: SYBKDK22  
 IBAN: DK02 8072 0001 0403 60

Klingenberg, 18/12/2024

sig. i. V. Helmut Schwarz  
Head of Customs and Export Control  
Corporate Logistics

This declaration has been printed automatically and has therefore no handwritten signature.  
However, it is valid as letter of commitment as per decree (EU) No. 2015/2447 art. 63 par. 3.

#####

**Export Restriction**

(1) The Customer/Recipient shall not, at any time, sell and/or export and/or re-export, directly or indirectly,  
a) to the Russian Federation; and/or  
b) to territories occupied by Russian Federation; and/or  
c) to the Republic of Belarus; and/or  
d) for use in the Russian Federation; and/or  
e) for use in territories occupied by Russian Federation; and/or  
f) for use in the Republic of Belarus  
directly or indirectly any  
i. item(s) herein listed ("Item(s)"); and/or  
ii. any intellectual property; and/or  
iii. know-how; and/or  
iv. any other confidential information whatsoever supplied or brought to attention to the Customer/Recipient howsoever under or in connection with this Agreement.

(2) The Customer/Recipient shall undertake its best efforts to ensure that the purpose of paragraph (1) is not frustrated by any third parties further down the commercial chain, including by possible resellers.

(3) The Customer/Recipient shall set up and maintain an adequate monitoring mechanism to detect conduct by any third parties further down the commercial chain, including by possible resellers, that would frustrate the purpose of paragraph (1).

(4) Any violation of paragraphs (1), (2) or (3) shall constitute a material breach of an essential element of this Agreement, and WIKAL shall be entitled to seek appropriate remedies, including, but not limited to:

- i. termination of this Agreement; and/or
- ii. a penalty of 10 % of the total value of this Agreement or price of the items exported, whichever is higher.

(5) The Customer/Recipient shall immediately inform WIKAL about any problems in applying paragraphs (1), (2) or (3), including any relevant activities by third parties that could frustrate the purpose of paragraph (1).  
The Customer/Recipient shall make available to WIKAL information concerning compliance with the obligations under paragraph (1), (2) and (3) within two weeks of the simple request of such information.

#####

Sales group                      Maria Stamm  
Phone                              +49 9372 132-8814  
E-Mail                              Maria.Stamm@wika.com